ETH zürich

Influence of head-over-body and body-over-head posture on craniospinal, vascular, and abdominal pressures in an acute ovine in-vivo model

Journal Article

Author(s):

Podgoršak, Anthony; Trimmel, Nina Eva; <u>Flürenbrock, Fabian</u> (b); Oertel, Markus Florian; Arras, Margarete; Weisskopf, Miriam; <u>Schmid Daners, Marianne</u> (b)

Publication date: 2023-08-02

Permanent link: https://doi.org/10.3929/ethz-b-000625123

Rights / license: Creative Commons Attribution 4.0 International

Originally published in: Fluids and Barriers of the CNS 20(1), <u>https://doi.org/10.1186/s12987-023-00458-9</u>

Funding acknowledgement: 184913 - Quantitative study on the pathophysiology of hydrocephalus (SNF)

RESEARCH

Open Access



Influence of head-over-body and body-over-head posture on craniospinal, vascular, and abdominal pressures in an acute ovine in-vivo model

Anthony Podgoršak¹, Nina Eva Trimmel², Fabian Flürenbrock¹, Markus Florian Oertel³, Margarete Arras², Miriam Weisskopf² and Marianne Schmid Daners^{1*}

Abstract

Introduction Optimal shunt-based hydrocephalus treatments are heavily influenced by dynamic pressure behaviors between proximal and distal ends of shunt catheters. Posture-dependent craniospinal, arterial, venous, and abdominal dynamics thereby play an essential role.

Methods An *in-vivo* ovine trial (n = 6) was conducted to evaluate communication between craniospinal, arterial, venous, and abdominal dynamics. Tilt-testing was performed between -13° and $+13^{\circ}$ at 10-min intervals starting and ending at 0° prone position. Mean pressure, pulse pressure, and Pearson correlation (*r*) to the respective angle were calculated. Correlations are defined as strong: $|r| \ge 0.7$, mild: 0.3 < |r| < 0.7, and weak: $|r| \le 0.3$. Transfer functions (TFs) between the arterial and adjacent compartments were derived.

Results Strong correlations were observed between posture and: mean carotid/femoral arterial (r = -0.97, r = -0.87), intracranial, intrathecal (r = -0.98, r = 0.94), jugular (r = -0.95), abdominal cranial, dorsal, caudal, and intravesical pressure (r = -0.83, r = 0.84, r = -0.73, r = 0.99) while mildly positive correlation exists between tilt and central venous pressure (r = 0.65). Only dorsal abdominal pulse pressure yielded a significant correlation to tilt (r = 0.21). TFs followed general lowpass behaviors with resonant peaks at 4.2 ± 0.4 and 11.5 ± 1.5 Hz followed by a mean roll-off of -15.9 ± 6.0 dB/ decade.

Conclusions Tilt-tests with multi-compartmental recordings help elucidate craniospinal, arterial, venous, and abdominal dynamics, which is essential to optimize shunt-based therapy. Results motivate hydrostatic influences on mean pressure, with all pressures correlating to posture, with little influence on pulse pressure. TF results quantify the craniospinal, arterial, venous, and abdominal compartments as compliant systems and help pave the road for better quantitative models of the interaction between the craniospinal and adjacent spaces.

Keywords Arterial blood pressure, Cerebrospinal fluid, Hydrocephalus, Intracranial pressure, Sheep model, Transfer function, Tilt testing

*Correspondence: Marianne Schmid Daners marischm@ethz.ch Full list of author information is available at the end of the article



© The Author(s) 2023. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit http://creativecommons.org/licenses/by/4.0/. The Creative Commons Public Domain Dedication waiver (http://creativecommons.org/publicdomain/zero/1.0/) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

Introduction

Hydrocephalus is characterized by disturbed cerebrospinal fluid (CSF) dynamics [1]. Its treatment has been a topic of debate ever since the first documented use of a valved CSF shunt 72 years ago [2]. Unfortunately, issues that limit the efficacy of shunts, namely subtle pressure fluctuations between the proximal and the distal end of the shunt system, which lead to frequent over- and under-drainage, are to this day not adequately addressed [3]. The ability to design more sophisticated treatment options is limited until gaps in quantitative understanding of CSF dynamics and their communication, in particular, the propagation of pressure from one compartment to another, are filled. Quantifying the interactions between craniospinal, arterial, venous, and intra-abdominal pressures during postural changes can help elucidate the dynamic relationships that exist between these systems, both improving our understanding of the underlying physiology that exists as well as pave the way for improved treatment options to be developed.

The influence of posture on physiological factors, such as intracranial pressure (ICP) [4], functional connectivity [5], cerebral blood flow [6], cerebral venous outflow [7], glymphatic transport of CSF [8], intracranial compliance [9], and cerebral perfusion pressure [10] has been well documented within the literature. In addition, specific investigations into the role of body position on cranial and lumbar CSF pressures in states of normal and impaired craniospinal communication have previously been reported in cats [11]. Furthermore, it is known that body position heavily influences arterial, venous, and respiratory function [12–14].

Previous studies have helped form the foundation of the common physiological understanding of the impact of posture on CSF and its interplay between other anatomical compartments [15–17]. Tilt-tests have been ubiquitously used in clinical and research settings since 1986 when the head-over-body tilt was first described as a non-invasive tool to investigate unexplained syncopes [18]. Exposure of a subject to controlled orthostatic changes in a safe, monitored, laboratory environment, specifically tilt tests are still used today to replicate symptoms and induce autonomic regulatory changes via manipulations to a body position that would otherwise be impossible [19, 20]. Dynamics in the CSF system are governed by the arterial [21, 22], venous [16], and abdominal [23] compartments. Therefore, it is important to acquire a comprehensive quantitative understanding of how these different compartments interact with each other in addition to how they influence intracranial and intrathecal pressures.

Quantifying reactions via mean pressures allows the investigation of bulk pressure reactions and communications within and between compartments [17, 24]. Furthermore, pulse pressure changes provide insights into physiological regulation and system compliances [24]. Transfer functions (TFs) are used to describe the input–output behavior of a system. Additionally, they can be used to model a system and investigate respective dynamics [25, 26]. TFs have previously been used to study dynamic relationships between CSF and cerebral blood flow [27, 28]. However, there exists a research gap in comprehensive studies simultaneously investigating the interplay of intracranial, intrathecal, arterial, venous, and abdominal systems in-vivo.

Thus, in this study, we measured physiological pressure changes of various anatomical compartments simultaneously during tilt testing of anesthetized sheep. The interrelationships are presented and quantified to provide insights into the physiological reactions under gravitationally induced dynamic changes.

Methods

Ethical statement

Animal housing and all experimental procedures were approved by the local Committee for Experimental Animal Research (Cantonal Veterinary Office Zurich, Switzerland) under the license number ZH119/2019, conforming to the European Directive 2010/63/EU of the European Parliament and the Council on the Protection of Animals used for Scientific Purposes, as well as to the Guide for the Care and Use of Laboratory Animals [29].

Anesthesia and animal instrumentation

Six adult female white alpine sheep (age 3.1 ± 1.2 yrs, Body Weight (BW) 75.6 ± 12.1 kg) were included in this study. Anesthesia was induced by intravenous injection of ketamine hydrochloride (Ketasol®-100 ad us. vet.; Dr. E. Graeub AG, Berne, Switzerland; 3 mg/ kg BW in combination with midazolam (Dormicum[®], Roche Pharma AG, Reinach, Switzerland; 0.2 mg/ kg BW and propofol (Propofol®- Lipuro 1%, B. Braun Medical AG; Sempach, Switzerland; 2-5 mg/ kg BW. After intubation, anesthesia was maintained by positive pressure ventilation (12-15 breaths/min, tidal volume 10–15 mL/kg, Fraction of Inspired Oxygen (FiO2) 0.5) of 2-3% isoflurane in oxygen/air mixture and a continuous infusion pump applying propofol (Propofol[®]- Lipuro 1%, B. Braun Medical AG; Sempach, Switzerland 2-4 mg/kg BW/h). Throughout the procedure, the animals additionally received a continuous intravenous infusion of sufentanil (Sufenta® Forte, Janssen-Cilag AG, Zug, Switzerland; 0.05 mg/kg BW/h). In all sheep, an arterial line (4 Fr) for the measurement of carotid arterial blood pressure (cABP) and a multilumen central venous catheter via the jugular vein for

the measurement of central venous pressure (CVP) were placed under percutaneous ultrasound guidance. Additionally, in four out of six sheep, a femoral arterial line (10F, Avanti[®], Cordis[®] Corporation, Miami Lakes, Florida, USA) for femoral arterial blood pressure (fABP) measurement and an 18G venous catheter in the proximal jugular vein to measure jugular venous pressure (JVP) were inserted. For measurement of ICP, a 9 Fr catheter (Ref. 55-3000, Neuromedex GmbH, Hamburg, Germany) was placed in the right lateral ventricle, confirmed by CSF egression, and fixed with Ethicon Bonewax (Johnson & Johnson Medical Ltd., Livingston, UK) to avoid CSF leakage. A 4.5 Fr Neuromedex catheter (Ref. 61-1400) was placed in the subdural space to measure intrathecal pressure (ITP) directly at the source via a laminotomy at level L6-7, also sealed with bone wax [30]. Hydrostatic equivalence was maintained between the intraventricular and intrathecal transducers by zeroing them to atmospheric pressure at the level of the lateral ventricles, both arterial and central venous sensors at the right atrial level, and the jugular venous and abdominal sensors at the location of the respective measurement source. Abdominal measurements were acquired directly at the specific abdominal locations via a peritoneal approach with 10 Fr access sheaths (Avanti, Cordis Corporation, Arrow Int. Inc., Reading, PA, USA) at the four different abdominal quadrants: cranial, caudal, dorsal, and ventral (IAPcr, IAPcd, IAPds, IAPve). Abdominal integrity was re-established via the closure of the peritoneum and fascia continuously and the skin via purse-string sutures. Intravesical pressure (IVP) was measured by connecting a transducer directly to the urinary catheter. All transducers were placed at the location of the respective measurement except for the carotid arterial, femoral arterial, and central venous pressures.

The sheep were then placed in the sternal position throughout the experiment, mimicking the horizontal position in humans at which the ITP and ICP were assumed not to be influenced by hydrostatic variations. Figure 1 shows all sensor locations and pressure signals acquired.

Experimental protocol

The experimental protocol in this study was designed to reveal how the dynamics between various physiological pressures (Fig. 1) change during standardized changes in body position. Laying on a surgical tilt table in sternal recumbence, body positions of 0° , $+5^{\circ}$, $+10^{\circ}$, $+13^{\circ}$, $+10^{\circ}$, $+5^{\circ}$, 0° , -5° , -10° , -13° , -10° , -5° , and 0° were applied; each step was maintained for 10-min. Positive angles denote Head-over-Body (HoB) posture while negative angles indicate Body-over-Head (BoH) posture. The tilt-table operated via a bi-directional electrical actuator with the motor just behind the caudal-most point on the sheep, leading to the fulcrum being around the midplane of the sheep to support proper weight distribution throughout the testing.

Data acquisition, pre-processing, and statistical analyses

All data was acquired using the commercially available software, Ponemah v5.1 (Data Science International, St. Paul, USA) with the ACQ-7700 acquisition unit using the Universal XE and ABCD 4 to amplify the signal. All data was acquired at a sampling frequency of 1 kHz, discriminated to 100 Hz for postprocessing retaining possible high frequency waveform components, and outliers were rejected via a z-score rejection method with a $\sigma_{\rm crit}$ of 3. Statistical significance of pressure changes was assessed via a paired dependent t-test with tilt angle defined as the independent variable and each pressure signal as dependent. Pressure differences between each step were first averaged across each subject (N=6) before being entered



- Carotid blood pressure (cABP)
 Femoral blood pressure (fABP)
- 3. Intracranial pressure (ICP)
- 4. Intrathecal pressure (ITP)
- 5. Central venous pressure (CVP)
- 6. Jugular venous pressure (JVP)
- 7a. Abdominal: cranial (IAPcr)
- 7b. Abdominal: dorsal (IAPds)
- 7c. Abdominal: caudal (IAPcd)
- 7d. Abdominal: ventral (IAPve)
- 8. Intravesical pressure (IVP)

Fig. 1 Pressure measurement locations during this study. The color scheme corresponds to the subsequent results

into the statistical model, yielding results representative of the entire cohort. Significance was determined at a level of p < 0.05. In addition, Pearson correlations are reported. Correlations are considered to be strong if $|\mathbf{r}| \ge 0.7$, mild if $0.3 < |\mathbf{r}| < 0.7$, and weak if $|\mathbf{r}| \le 0.3$. Regression analysis was performed considering changes in pressure across the entire sheep cohort from the initial mean baseline with all sheep having a complete dataset with no rejections. All analyses were performed using custom scripts in Python 3.7.10 (Open Source, Python Software Foundation, Wilmington, Delaware, USA). The data is available open access at the following link: https://doi.org/10.3929/ethz-b-000623698.

4th order forward/backward Butterworth filter with a cut-off frequency of 0.1 Hz [10]. Mean pressures were defined as 10-min arithmetic means over the entire tilt step once pressures had stabilized (Fig. 2). Note that cABP, fABP, and IAPve change slightly during the initial baseline 0° period: these are normal pressure fluctuations and are considered during the calculation of mean pressure changes.

Pulse pressure

For all signals, reactions in pulse pressure were calculated as the difference between peak and trough (Fig. 3). Pulse pressures for each waveform were calculated and then averaged over the entire length of the tilting step.

Mean pressure

Mean pressure reactions to different body positions were calculated after the data was pre-processed as follows. To remove effects of the cardiac and respiratory waveforms, the data was lowpass filtered using a

Transfer functions

Transfer functions (TFs) were calculated by dividing the output and the input signal in the frequency domain. As the heart governs all vascular pressures in the body, cABP was defined as the input and all other pressure



Fig. 2 Visualization of mean pressure during body-over-head (BoH) positions in a single sheep; the respective steps are indicated in the range of 0° to – 13° to 0° on top of the figure. Note that head-over-body (HoB) positions are not shown on this figure but were considered in all analyses and that elapsed time is shown on the x-axis. cABP: carotid arterial blood pressure; fABP: femoral arterial blood pressure; ICP: intracranial pressure; ITP: intrathecal pressure; CVP: central venous pressure; JVP: jugular venous pressure; IAPds: dorsal intra-abdominal pressure; IVP: intravesical pressure; IAPve: ventral intra-abdominal pressure; IAPcr: cranial intra-abdominal pressure; IAPcd: caudal intra-abdominal pressure



Fig. 3 Calculation of pulse pressure. On the right-hand side, the respective measurement (from maximum to minimum) is indicated numerically. cABP_{amp}: carotid arterial blood pressure amplitude; fABP_{amp}: femoral arterial blood pressure amplitude; ICP_{amp}: intracranial pressure amplitude; ITP_{amp}: intrathecal pressure amplitude; CVP_{amp}: central venous pressure amplitude; JVP_{amp}: jugular venous pressure amplitude; IAPds_{amp}: dorsal intra-abdominal pressure amplitude; IAPcd_{amp}: caudal intra-abdominal pressure amplitude; IAPcd_{amp}; caudal intra-

signals, such as ICP, ITP, or CVP as the individual outputs. The conversion between time-series to frequency domain was done via discrete fast Fourier transformation (FFT) over the 0° , +13°, and –13° steps with a rectangular 10-min window without overlap. Each TF was then smoothed using Gustaffson's method to avoid transients at the edges of the signal while retaining trends [31]. These results were then averaged across all six sheep to represent our cohort's final TFs. A 95% confidence interval was plotted in addition to the averaged TF. The full 1 kHz data resolution was used in the derivation of each TF to ensure the highest spectral resolution possible.

Results

Mean reactions

cABP and fABP show negative correlations with tilt angles, i.e., as tilt angle increases, mean pressure decreases (Table 1). ICP and ITP (horizontal anatomic distance from ITP to ICP sensor locations of 70.0 ± 3.2 cm, respective vertical distance of 0.4 ± 0.8 cm) reacted

opposingly to each other, highlighting their hydraulic connection via the spinal canal. Minimal reaction and mildly positive correlation in CVP were observed while JVP revealed a similar strongly negative correlation with tilt angle as cABP and fABP. There were less obvious patterns with IAPs, however a mild direct relationship with tilt angle can be observed in IAPcd (horizontal anatomic distance from IAPcd to ICP sensor locations of 60.5 \pm 4.3 cm, respective vertical distance of 9.5 \pm 5.5 cm) and IVP (horizontal anatomic distance from IVP to ICP sensor locations of 76.4 \pm 5.2 cm, respective vertical distance of 16.8 ± 3.6 cm) and a mild indirect relationship in IAPcr (horizontal anatomic distance from IAPcr to ICP sensor locations of 56.7 \pm 5.6 cm, respective vertical distance of 0.9 ± 2.6 cm), IAPds (horizontal anatomic distance from IAPds to ICP sensor locations of 61.3 ± 4.7 cm, respective vertical distance of -3.7 ± 2.7 cm), and IAPve (horizontal anatomic distance from IAPve to ICP sensor locations of 62.7 ± 4.8 cm, respective vertical distance of 7.3 ± 3.5 cm) (Fig. 4). Equal and opposite

Table 1	Mean pressur	re changes ov	er the six shee	p from the first	baseline pron	e position (0 ₁ °) listed first as	averaged valu	ues then as th	e respective rel	ative values fi	om 0 ₁
Angle (°)		cABP (mmHg)	fABP (mmHg)	ICP (mmHg)	ITP (mmHg)	CVP (mmHg)	JVP (mmHg)	IAPcr (mmHg)	IAPcd (mmHg)	IAPds (mmHg)	IAPve (mmHg)	IVP (mmHg)
01	Mean Pres- sure	80.4±16.2	75.5±17.7	16.8±9.7	18.2±8.2	7.6±4.0	5.1±1.6	15.9±3.2	14.9±1.7	12.6±0.8	10.0±0.4	11.4±3.2
+13	\bigtriangledown	-8.6 ± 5.8	-4.6 ± 5.8	-5.9 ± 3.3	$+5.0\pm3.2$	+0.4±4.3	-1.7 ± 3.1	-4.7±3.1	+2.4±7.4	-6.3 ± 18.9	-2.6 ± 6.2	$+7.0 \pm 1.9$
02	\bigtriangledown	-2.4 ± 6.2	-0.3 ± 2.5	$+0.9\pm3.7$	-1.3 ± 3.4	-0.1 ± 1.9	$+ 0.0 \pm 0.6$	-3.2 ± 6.8	$+2.4 \pm 6.8$	-3.7 ± 10.5	+1.4±4.5	$+0.8 \pm 2.1$
-13	\bigtriangledown	$+ 6.1 \pm 8.7$	$+1.5 \pm 7.3$	$+10.8\pm 5.6$	-3.4±4.9	-0.1 ± 1.6	+ 2.9±1.9	$+6.4 \pm 20.6$	-0.8 ± 8.4	-1.9 ± 22.3	-1.3 ± 2.2	-6.1 ± 2.9
03	\bigtriangledown	-1.5 ± 6.0	-1.0 ± 5.4	+4.0±5.2	$+1.3 \pm 3.9$	-0.1 ± 2.6	-0.9 ± 1.9	-1.7±12.5	$+0.4 \pm 4.1$	-4.9 ± 22.8	-2.6 ± 2.8	-1.3 ± 2.3
r		- 0.97	- 0.87	-0.98	0.94	0.65	- 0.95	-0.83	0.84	-0.73	-0.19	66.0
<i>P</i> -value		.036	.014	.030	.036	.045	.047	.044	0.032	.014	.030	180.
Angle (°)		cABP _{amp} (mmHg)	fABP _{amp} (mmHg)	ICP _{amp} (mmHg)	ITP _{amp} (mmHg)	CVP _{amp} (mmHg)	JVP _{amp} (mmHg)	IAPcr _{amp} (mmHg)	IAPcd _{amp} (mmHg)	IAPds _{amp} (mmHg)	IAPve _{amp} (mmHg)	IVP _{amp} (mmHg)
01	Mean Ampli- tude	18.8±6.6	26.0±9.6	2.6±1.2	3.0±1.1	3.7±2.0	1.2 ± 0.8	7.7±7.6	4.6 ±3.9	11.±10.2	6.2 ± 7.2	1.7 ± 0.8
+ 13	\bigtriangledown	-0.7 ± 5.2	-3.2 ± 6.1	-0.1 ± 0.6	-0.4±1.1	$+0.2 \pm 1.7$	$-0.5 \pm .05$	-5.1 ± 8.0	-1.8 ± 7.1	-9.0 ± 10.9	-3.6 ± 7.8	$+ 0.0 \pm 0.2$
02	\bigtriangledown	$+3.0\pm2.5$	-1.9 ± 4.7	$+0.4 \pm 0.8$	-0.5 ± 0.8	$+1.1 \pm 1.6$	$+ 0.2 \pm 0.3$	$+ 0.7 \pm 5.1$	-3.2 ± 6.9	-2.8 ± 18.7	-1.7±4.9	$+ 0.0 \pm 0.4$
-13	\bigtriangledown	$+0.3 \pm 8.2$	-1.5 ± 5.1	$+0.9\pm0.9$	-0.3 ± 0.7	$+0.1 \pm 1.2$	$+ 0.9 \pm 1.3$	-6.2 ± 9.8	-2.4 ± 7.0	-7.9 ± 9.6	-2.1 ± 3.8	-0.3 ± 0.4
03	\bigtriangledown	$+ 2.1 \pm 2.3$	- 3.9±4.8	$+0.7 \pm 1.1$	$+0.0 \pm 0.8$	$+0.3 \pm 1.0$	$+ 0.5 \pm 0.8$	-4.4 ± 6.7	-2.7 ± 6.5	-10.5 ± 10.8	-3.4 ± 7.1	-0.6 ± 0.7
r		- 0.49	0.31	-0.91	0.16	-0.07	- 0.93	0.15	0.37	0.21	- 0.01	0.66
<i>P</i> -value		.292	.232	.434	.465	.488	.459	.217	.168	.025	.183	.471
(+ 13°, 02°, cranial intr for cABP _{am} amplitude; IAPve _{amp} : v to the first	 – 13°, 03°) for cAE a-abdominal presign carotid arterial yP_{amp}; jugular v entral intra-abdo baseline (0, °) 	SP: carotid arteria ssure; IAPcd: cauc blood pressure a enous pressure a minal pressure; a	il blood pressure; dal intra-abdomin implitude; fABP ^{am} implitude: IAPCr _{ar} and IVP _{amp} : intrave	fABP: femoral arteri. al pressure; IAPds: c p: femoral arterial b ^{p:} cranial intra-abd ^{sic} cranial pressure. All v	al blood pressure dorsal intra-abdor lood pressure am iominal pressure. 'alues are in mm ⁺	; ICP: intracranial minal pressure; IA pilitude; ICP _{amp} : in amplitude; IAPcd, ig (mean ± SD) wi	pressure; ITP: intr Pve: ventral intra utracranial pressu amp: caudal intra-e th Pearson correl.	athecal pressure -abdominal pres re amplitude; ITF abdominal press ation (r) and p-v	; CVP: central ven sure; and IVP: intr $_{amp}$: intrathecal $_{amp}$: ure amplitude; IA alues reported. Δ	nous pressure; JVP: ravesical pressure ? pressure amplitude Pds _{amp} : dorsal intrr denotes a change	jugular venous _i as well as mean _f a: CVP _{amp} : central a-abdominal pre in pressure from	ressure; for IAPcr: oulse pressures venous pressure ssure amplitude; the angle relative

_	
Q	
Ø	
Ę	
<u>a</u>	
<u>ع</u>	
Š	
Ū	
e S	
SS	
2	
Ę	
st	
a	
- L	
÷	
ŝ	
Ď	
م م	
Q	
g	7
D D	
Š	
o:	
a	
St	
Ļ	
0	
ste	
· <u> </u>	
ି	`
9	,
5	
Ę	
S	
ă	
e	
ō	
prol	
ne proi	
eline proi	
iseline proi	
baseline proi	
st baseline proi	
first baseline proi	
ne first baseline proi	
the first baseline proi	
m the first baseline proi	
from the first baseline pro	
p from the first baseline pro	
sep from the first baseline proi	
sheep from the first baseline pro	
x sheep from the first baseline proi	
six sheep from the first baseline proi	
he six sheep from the first baseline pro	
r the six sheep from the first baseline pro	
ver the six sheep from the first baseline pro	
over the six sheep from the first baseline pro	
es over the six sheep from the first baseline pro	
laes over the six sheep from the first baseline pro	
anges over the six sheep from the first baseline pro	
changes over the six sheep from the first baseline pro	
re changes over the six sheep from the first baseline pro	
ure changes over the six sheep from the first baseline pro	
sssure changes over the six sheep from the first baseline pro	
oressure changes over the six sheep from the first baseline pro	
n pressure changes over the six sheep from the first baseline pro	
aan pressure changes over the six sheep from the first baseline pro	
Mean pressure changes over the six sheep from the first baseline pro	



Fig. 4 Visualization of mean changes across the entire range of tilt angles and all pressures considered (N=6). Error bars indicate one standard deviation at each tilt angle. cABP: carotid arterial blood pressure; fABP: femoral arterial blood pressure; ICP: intracranial pressure; ITP: intrathecal pressure; CVP: central venous pressure; JVP: jugular venous pressure; IAPcr: cranial intra-abdominal pressure; IAPcd: caudal intra-abdominal pressure; IAPds: dorsal intra-abdominal pressure; IVP: intravesical pressure

correlations were observed in IAPcr and IAPcd, while IVP yielded a strong correlation to tilt (r=0.99). A positive vertical distance denotes when the respective sensor transducer is below the ICP transducer, and negative when above the ICP transducer.

Table 1 shows the relative changes in mean pressures at 0° baseline prone position (0_1°) and the maximum angle $(+13^{\circ})$ and minimum angle (-13°) . Additionally, relative changes between baseline (0_1°) second prone 0_2° and third prone position 0_3° are shown. Similar behaviors in cABP and fABP were observed, where both decrease at $+13^{\circ}$, followed by an increase at -13° . cABP and fABP remained decreased upon the return to 0° (0_{2}° and 0_{3}°) compared to the first 0° baseline 0_{1}° . ICP and ITP as well as IAPcr and IAPcd showed opposing reactions to each other i.e., as IAPcr increased, IAPcd decreased, highlighting their 180° measurement directions. IAPds and IAPve did not show opposing behavior as observed in the cranial and caudal measurements. IVP showed similar behavior to IAPcd. IAPcr, IAPcd, and IVP all returned to values within one SD of the original baseline while IAPds and IAPve were considerably below their original baseline pressure value. A large standard deviation was observed across the entire range of tilt angles in IAPds.

All changes in mean proved to be statistically significant, with cABP, ICP, and JVP, IAPcr, and IAPds all being strongly negatively correlated to tilt. CVP proved weakly positively correlated, and ITP being strongly positively correlated with tilting angle. IAPcd and IVP proved to be strongly positively correlated with tilting angle. In all pressures, disproportionate reactions to tilt angle were observed, especially in ICP, with $+13^{\circ}$ yielding a pressure change of -5.9 mmHg while -13° yielded a pressure change of 10.8 mmHg compared to the initial baseline.

Pulse pressure reactions

Less noticeable patterns and weaker correlations were observed in pulse pressure amplitudes (Fig. 5). A slight increase in cABP_{amp} from 0° to + 13° corresponded with a decrease in fABP_{amp} over the same time window. There were slight changes in ICP_{amp} and ITP_{amp} over the range of tilt angles, however they lacked significance. CVP_{amp} and JVP_{amp} remained relatively constant across the entire range. From 0° to + 5°, there was an increase in IAPcr_{amp} and IAPds_{amp} amplitude with a decrease of IAPcd_{amp} and IAPve_{amp}. However, for the remainder of the tilt test, all IAPs followed a similar pattern between each other while IVP remained relatively constant across the entire range. Only changes seen in IAPds_{amp} proved to be statistically significant.



Fig. 5 Visualization of pulse pressure changes across the entire range of tilt angles and all pressures considered (N=6). Error bars indicate one standard deviation at each tilt angle. cABP, carotid arterial blood pressure; fABP, femoral arterial blood pressure; ICP, intracranial pressure; ITP: intrathecal pressure; CVP: central venous pressure; JVP jugular venous pressure; IAPcr: cranial intra-abdominal pressure; IAPcd: caudal intra-abdominal pressure; IAPds: dorsal intra-abdominal pressure; IAPve: ventral intra-abdominal pressure; IVP: intravesical pressure

Transfer functions

Dynamic transmission patterns were observed across all TFs with considerably different magnitudes across 0° , +13°, and –13° (Fig. 6). At 0° in the arterio-craniospinal, arterio-venous, and arterio-abdominal TFs, there exist an initial soft decay; however, this observation is not as strong in the +13° and –13° TFs. Following this initial behavior there exist dynamic resonant peaks at 4.2 ± 0.4 and 11.5 ± 1.5 Hz before a final decay with roll-offs of – 14.9 ± 5.3 , -17.0 ± 4.3 , -17.0 ± 3.2 , -15.1 ± 6.3 , -14.3 ± 7 . 2, -16.2 ± 6.5 , -14.9 ± 5.5 , and -17.7 ± 2.9 dB/decade for cABP-ICP, cABP-JVP, cABP-IAPcr, cABP-IAPcd, cABP-IAPds, cABP-IAPve, and cABP-IVP, respectively. These local minima and maxima are more prevalent in the arterio-venous and arterio-CSF TFs when compared to the arterio-abdominal TFs.

Discussion

Pressure responses

Across the entirety of the tilt tests, ICP and ITP revealed opposite reactions to the orthostatic changes. ICP was in a strong indirect relationship with the tilt angle causing a decrease in mean ICP as the angle became more HoB. This hydrostatic change may have caused pressure to drop via a gravitationally induced relaxation of the cerebrovascular bed and an increase in compliance. Conversely, as the sheep becomes more BoH, CSF and venous blood pool in the cranium, causing an increase in their mean pressure. Similar volumetrically-induced changes in pressure dynamics have previously been reported [21]. Interestingly, ICP and ITP did not return to their pretilt baseline, ending slightly higher. This may indicate an autonomic compensation via the orthostatic stressing of the CSF [4] and glymphatic system leading to a temporary reduction in CSF absorption, causing increased pressure post-tilt test. The order of the tilt test (i.e., starting with BoH rather than HoB) may further impact the final return to baseline. The influences of these different positions are split between gravitational influences and the resultant volumetric pressure changes. However, intracranial and lumbar intrathecal spaces have similar relaxation times when a volumetric pressure change is introduced [24]. Furthermore, recent work has shown that there exists similar venous dynamics in sheep and humans in response to tilt [32, 33], therefore there may be a level of venous collapse that effectively reduces CSF outflow during upright postures, maintaining ICP and reducing the influence of HoB posture, as also supported by the influence as dictated by the ICP-ITP hydrostatic column (15.7 cm at \pm 13°, calculated using trigonometry, leading to 11.6 mmHg) being suggesting a larger pressure change than what was observed. Therefore, these findings suggests that gravitational influences may cause uneven dynamic changes in HoB and BoH postures.

Both cABP and fABP responded similarly to the orthostatic influence on ICP, but most likely induced



Fig. 6 Transfer functions (TF) of all measured pressures averaged over all sheep (N=6). Blue, initial prone position 0₁°; red, + 13°; and green, - 13° tilt angles. A 95% confidence interval is plotted following the same color scheme. All TFs are computed with the carotid arterial blood pressure (cABP) as the input to each individual output: ICP: intracranial pressure; ITP: intrathecal pressure; CVP: central venous pressure; JVP: jugular venous pressure; fABP: femoral arterial blood pressure, IAPcr: cranial intra-abdominal pressure; IAPcd: caudal intra-abdominal pressure; IAPds: dorsal intra-abdominal pressure; IAPve: ventral intra-abdominal pressure; and IVP: intravesical pressure

via different mechanisms. Vasovagal and baroreflex responses have been well described in humans to maintain stable hemodynamics following a change in body position [34, 35, 35, 36, 36, 37]. Furthermore, baroreceptor activity and venous dynamics when compared between newborn sheep and humans have been shown to be functionally similar [33, 38]. As gravity's directional influence on the cardiovascular system changes, baroreceptors create a negative feedback loop to maintain homeostasis and regulate the vasovagal syncope caused by tilting. Sato et al. [39] reported in humans that baroreflex is heavily attenuated under propofol-based anesthesia, which may motivate the apparent posture-induced changes to cABP and fABP in our sheep. There was minimal change in CVP across the range of tilts. As CVP dynamics are usually dictated by cardiopulmonary blood volume, perhaps the $\pm 13^{\circ}$ extremities were insufficient to provide this volume change potentially due to its location near its respective hydrostatic indifference point. JVP, however, displayed a clear indirect relationship with tilt angle similar to the arterial pressures and ICP. In humans, a split jugular venous system exists comprised of the internal and external jugular veins connected to the brachiocephalic vein. The internal jugular vein is thereby described to collapse during upright posture to regulate ICP via an increase in venous outflow resistance [16, 40]. This, in turn, causes the vertebral venous plexus to act as the primary venous outflow tract while upright [41]. However, in sheep there lacks an internal jugular vein, potentially leading to the vertebral plexus playing a more important role in physiological compensation to orthostatic variations than in humans, causing the more muted response in JVP due to relatively more venous drainage occurring in the vertebral veins.

The simultaneous measurement of IAP while undergoing tilting is less described in the literature, with only a few studies reporting changes in compartmental IAP following trauma and liver transplantation [23, 42, 43]. Our study motivates the idea that the intra-abdominal space is heterogeneous and should not be considered a unified entity. IVP is commonly used as a non-invasive indirect measurement of IAP in patients. Reasonable agreement, however, was only found between IAPcd and IVP in our sheep, thereby questioning the validity of IVP particularly in light of VP shunt placement and the flexibility and mobility of the distal shunt catheter, which is commonly implanted via an incision in the upper abdomen of patients, more cranially than the caudal sensor placement in our sheep cohort. A significant variation in IAP measured in the upper and lower abdominal compartment following liver transplantation in humans aggravated by postural changes has previously been described by Cresswell et al. [43]. We found IAPcr and IAPds to have negative correlations with tilt angle while IAPcd, IAPve, and IVP were observed to be positively correlated. Sheep, like other ruminants, have a highly asymmetric abdominal compartment, with the rumen occupying nearly all of the left side [44]. It might therefore be assumed that the regional IAP effect may not be as prominent in humans. However, this makes sheep unique in the development and testing of novel shunts, as potentially varying pressure gradients between proximal and distal ends of the shunt can be studied to improve shunt function. Furthermore, these ovine results motivate a deeper investigation into the concept of a unidimensional IAP measurement as it may fail to describe true IAP dynamics adequately.

These results, while showing distinct dynamics between the proximal (craniospinal) and distal (arterial, venous, or abdominal) ends of any shunt catheter, emphasize the correspondent paired behavior in the craniospinal, arterial, and venous compartments while also visualizing the heterogeneity of the abdominal space.

Pulse pressure responses

Whenever the CSF space and surrounding area endure a volumetric fluid increase, a reduction in compliance occurs due to the stretching of the cavity walls as a reaction to the increase in pressure, e.g. brain compliance is inversely proportional to ICP [45]. As CSF is dynamically pulsatile, the fluid's pulse pressure will also change as the cavity becomes less elastic [46]. In our study, gravitational forces acting upon the sheep lead to flooding of the venous bed and cerebral intraventricular space (in the case of BoH) and potentially a rush of CSF pooling in the lumbar regions (in the case of HoB). During BoH, cranial pooling of CSF causes the surrounding ventricular tissue to experience an acute reduction in compliance, leading to an increase in $\mathrm{ICP}_{\mathrm{amp}}$ and a corresponding decrease in ITP_{amp}. Oppositely, during the HoB position, CSF pools in the lumbar regions, causing an increase in ITP_{amp} and a decrease in ICP_{amp}, possibly following a reduction of lumbar compliance. As CVP remained relatively constant during tilting the pulse pressure also remained relatively stable. In JVP, a large dynamic change in mean pressure was observed, with no corresponding increase in pulse pressure. As the jugular vein is a high-compliance vessel, it is possible that the extent of tilting may not have induced enough of a volumetric change to effectively cause a reduction in compliance, leading to increased pulse pressures.

No obvious patterns were observed in the arterial pulse pressures across tilt angles, which may be a manifestation of $our - 13^{\circ}$ to $+ 13^{\circ}$ tilt steps. Furthermore, the arterial tree is a muscular system, which can counteract gravitational pooling, in contrast to the CSF and venous systems. This effect suggests a certain amount of independence

between ABP_{amp} and posture. No distinct pulse pressure patterns were observed in the abdominal pressures, some behavior even disagreed with the changes in mean pressure e.g., large changes in pulse pressure that are yet to have a physiological explanation. This may be attributed to the heterogeneity of the sheep's abdominal cavity. The ovine rumen is filled very specifically with fluid and solid matter on the bottom, followed by a layer of solid matter in the middle and gas on the very top. It is possible that the tilting of the sheep changed the layering of the ruminal content and caused the position of the abdominal organs to shift-both of which may have led to the seemingly spontaneous changes in pulse pressure. Overall, these results illustrate the fact that there exists minimal marked impact from tilting on pulse pressures in the craniospinal, arterial, venous, and abdominal spaces.

Transfer function analysis

The classical view of large-artery compliance is that it supports the dampening of strong cardiac ventricular ejections and assists in the reduction of the pulsatile nature of the flow into a more constant downstream flow at the site of the arterioles, thus supporting organ perfusion and ensuring consistent and manageable pulsatility to the downstream vessels [47]. In a similar physiological phenomenon, intracranial compliance has been described as essential to the compensatory mechanisms to maintain ICP stability and intracranial homeostasis [48].

In a mathematical sense, compliant mechanisms operate by attenuating higher frequencies-i.e., they act as lowpass filters. The general lowpass behavior of our derived TFs motivates the physiological doctrine of arteries, veins, the craniospinal space, and the abdomen behaving as compliant compartments. Furthermore, multiple resonant peaks (local maxima) exist among each TF, such as the cABP-ICP, cABP-ITP, or cABPfABP TF. The existence of complex yet correspondent resonant peaks in the arterio-craniospinal and arterio-venous TFs suggests that these system responses exhibit common features due to resonance, agreeing with the work of Wagshul et al. [49] yet disagreeing with the work of Tenti et al. [50], who argued that resonance plays no role in the synchronicity of arterial and CSF pulsations.

Interestingly, there exists behavior indicative of a notch filter at 8.0 Hz in cABP-IAPds and to a lesser extent in cABP-IAPve and cABP-IVP. Notch behavior has previously been reported [49], however it was only reported in cABP-ICP TFs. The TFs remained quantitatively similar across the three tilt angles $(0^\circ, +13^\circ, -13^\circ)$ with few exceptions. cABP-IAPcd and cABP-IAPds TF magnitudes at 0° remained asymmetrically larger

than their + 13° and - 13° counterparts until around 9 Hz. + 13° and - 13° were the two tilt extrema considered in this study and IAPcd and IAPds were facing the tail and back of the sheep, respectively. As introduced prior, the ovine rumen is comprised of transient layers. If an asymmetric shift in the rumen content occurred, this might have manifested itself in a similar asymmetric manner to the different abdominal pressure measurements. This could, in turn, have caused a corresponding asymmetric shift in the individual TFs between 0° , + 13°, and - 13°, which may also explain the cABP-IVP TF remaining asymmetrically larger than the $+13^{\circ}$ and -13° ones. Each of our TFs final decays are within the range of 20 dB/decade, which may lead to the assumption that first-order lowpass behavior dominates. However, bands of local maxima and minima (i.e., resonance) before the final decay suggest that some intercompartmental interactions are subject to higher order dynamics, possibly due to the existence of multiple physiological pathways from the carotid artery to the craniospinal, venous, and abdominal spaces. Nevertheless, these results show that we have similar transmission patterns across the investigated compartmental pairs and that there does exist attenuation of higher frequencies. This supports the concept that the arterial, craniospinal, venous, and abdominal systems serve as compliant mechanisms.

The derivation of TFs between the carotid arterial and adjacent compartments has revealed complex compliant systems which are dictated by multiple overlain higherorder dynamics.

Clinical and physiological implications

The optimal strategy of shunt-based treatments for hydrocephalus requires a detailed understanding of the dynamic pressure behaviors that exist between proximal and distal ends of the shunt catheters (i.e., intraventricular and peritoneal abdominal spaces for ventriculoperitoneal shunts, intraventricular and venous spaces for ventriculo-venous shunts, and intraventricular and atrial cardiac spaces for ventriculo-atrial shunts) [51]. This study provides a foundation to build this understanding via analyzing the dynamic reactions to tilting. Observing the impact of tilt angle on physiological pulse pressures is an important consideration when designing the control strategy behind any active shunts-specifically to avoid over- or under-drainage at the systolic peak or diastolic trough. This study shows that moderate HoB and BoH tilt have minimal impact on pulse pressures, meaning that, at least from -13° to $+13^{\circ}$ of postural change, there need not exist unique sophistications to adjust for changes in pulse pressure. However, an expansion of this investigation to the entire range of normal body postures (0° to

90°) is required. There was no indication of reflex tachycardia during HoB positions, which does not necessarily suggest that the reflex pathway is not present, rather perhaps that our moderate tilt angles were insufficient to trigger them. Moreover, the effect of posture on glymphatic drainage has also been a topic of interest in the field, with one study conducted in rats showing that the lateral sleeping position led to the most optimal glymphatic drainage when compared to the supine and right lateral decubitus positions [52]. It could also be the case that, depending on the brain's position relative to the fulcrum of the tilt, glymphatic drainage may vary. Furthermore, the compliant nature of most body parts has been a topic of considerable discussion ever since Spencer and Denison [53] first quantified arterial compliance in 1963. The derived TFs have quantified the intercompartmental compliant behavior, revealing which specific frequencies are attenuated to what extent. Moreover, having the TFs between arterial, venous, craniospinal, and abdominal compartments supports the investigation into specific behaviors at all possible spectral inputs, paving the way for a more quantitative understanding of how compliance plays a role in physiology.

Physiological considerations

This investigation was performed only in female sheep; however, no sex-based pressure differences are anticipated. A balanced, multimodal anesthesia was used, with the intention to reduce the dose of each anesthetic agent and thus their side effects. General anesthesia is known to affect autoregulatory mechanisms to a certain extent, which is common limitation in physiological studies performed under general anesthesia. However, due to the nature of manipulations performed in this experiment, general anesthesia was necessary and was considered in the interpretation of all findings. Moreover, our acute model under anesthesia ensures no nodding of the head during movements, effectively avoiding head-movementinduced hydrostatic influences.

Limitations

The use of mechanical positive pressure ventilation could not only have had an impact on venous and thoracic pressure but also on the flow of CSF, which may have influenced our results. Furthermore, as sheep are quadrupeds, elements specific to their physiological fluid dynamics limit the transferability to humans. Specifically, quadrupeds are physiologically adapted to changing CSF axis closer to the neck when compared to humans where the CSF axis is firstly vertical most of the time, and secondly much more prone to larger changes. This may also lead to limitations in translating CSF-specific relationships to humans. The ovine abdominal space is anatomically different from that of a human, with multiple different layers comprised of different contents and ruminal filling stages. The effects suspected to have been caused by regional abdominal heterogeneities observed in this study may not be as distinct in humans, whose abdominal anatomy is different and not as complex as the ovine system. The tilt-table used in this study was limited to $\pm 13^{\circ}$ in a prone position, which might limit conclusions and direct comparisons to human body positions, such as standing, sitting, or supine.

Conclusions

Our investigations have provided comprehensive insights into the dynamic physiological relationships between the pressure conditions at the proximal and distal positions of any shunt catheter during changes in body position. Orthostatic pressure changes from changes in body position have been shown to significantly influence mean pressures in sheep, while minimal patterns in pulse pressure suggest potential independence between body position and pulse, at least at moderate tilts in posture. These relationships suggest that pulse pressure may be considered as constant and could be exploited as control parameter. Transfer function analyses quantify previously unquantified physiological insights: there exists a general compliant, lowpass theme to the arterio-craniospinal, arterio-venous, and arterio-abdominal systems. However, there are also overlain higher-order dynamics present.

Acknowledgements

The authors acknowledge Dr. Lennart Stieglitz (Department of Neurosurgery, University Hospital Zurich) for his clinical insight regarding the protocol for this study. This work is affiliated with the Clinical Neuroscience Center, University Hospital Zurich.

Author contributions

Conception and design: MSD, MW, MA, AP, NET, MFO. Acquisition of data: MSD, MW, AP, NET. Analysis and interpretation of data: all authors. Drafting the article: AP. Critically revising the article: all authors. Study supervision: MSD, MA, MW.

Funding

Open access funding provided by Swiss Federal Institute of Technology Zurich. The authors acknowledge funding from the Swiss National Science Foundation through Grant 315230_184913 and the Innosuisse project 42749_1 IP-LS.

Data availability

All relevant in-vivo data used in this study can be found open access using the following link: https://doi.org/10.3929/ethz-b-000623698.

Declarations

Competing interests

The authors of the manuscript titled "Influence of head-over-body and bodyover-head posture on craniospinal, vascular, and abdominal pressures in an acute ovine in-vivo model" report no competing interests concerning the materials or methods used in the study or the findings specified in this paper.

Author details

¹Department of Mechanical and Process Engineering, ETH Zurich, Zurich, Switzerland, ²Center for Surgical Research, University Hospital Zurich, University of Zurich, Zurich, Switzerland, ³Department of Neurosurgery, University Hospital Zurich, University of Zurich, Zurich, Switzerland.

Received: 25 January 2023 Accepted: 13 July 2023 Published online: 02 August 2023

References

- 1. Rekate HL. The definition and classification of hydrocephalus: a personal recommendation to stimulate debate. Cerebrospinal Fluid Res. 2008;5:2.
- 2. Nulsen FE, Spitz EB. Treatment of hydrocephalus by direct shunt from ventricle to jugular vain. Surg Forum. 1951;399–403.
- Gehlen M, Eklund A, Kurtcuoglu V, Malm J, Schmid DM. Comparison of anti-siphon devices—how do they affect CSF dynamics in supine and upright posture? Acta Neurochir (Wien). 2017;159(8):1389–97.
- Qvarlander S, Sundström N, Malm J, Eklund A. Postural effects on intracranial pressure: modeling and clinical evaluation. J Appl Physiol. 2013;115:1474–80.
- Avelar-Pereira B, Tam GKY, Hosseini SMH. The effect of body posture on resting-state functional connectivity. Brain Connect. 2022;12(3):275–84.
- Alperin N, Lee SH, Sivaramakrishnan A, Hushek SG. Quantifying the effect of posture on intracranial physiology in humans by MRI flow studies. J Magn Reson Imaging. 2005;22(5):591–6.
- Sagirov AF, Sergeev TV, Shabrov AV, Yurov AY, Guseva NL, Agapova EA. Postural influence on intracranial fluid dynamics: an overview. J Physiol Anthropol. 2023;42(1):5. https://doi.org/10.1186/s40101-023-00323-6.
- Bohr T, Hjorth PG, Holst SC, Hrabětová S, Kiviniemi V, Lilius T, et al. The glymphatic system: current understanding and modeling. iScience. 2022;25(9):104987.
- D'Antona L, Craven CL, Bremner F, Matharu MS, Thorne L, Watkins LD, et al. Effect of position on intracranial pressure and compliance: a crosssectional study including 101 patients. J Neurosurg. 2022;136(6):1781–9.
- Petersen LG, Petersen JCG, Andresen M, Secher NH, Juhler M. Postural influence on intracranial and cerebral perfusion pressure in ambulatory neurosurgical patients. Am J Physiol Regul Integr Comp Physiol. 2016;310(1):R100–4.
- Klarica M, Radoš M, Erceg G, Petošić A, Jurjević I, Orešković D. The influence of body position on cerebrospinal fluid pressure gradient and movement in cats with normal and impaired craniospinal communication. PLoS ONE. 2014;9(4): e95229.
- Katz S, Arish N, Rokach A, Zaltzman Y, Marcus EL. The effect of body position on pulmonary function: a systematic review. BMC Pulm Med. 2018;18(1):1–16.
- Trinity JD, McDaniel J, Venturelli M, Fjeldstad AS, Ives SJ, Witman MAH, et al. Impact of body position on central and peripheral hemodynamic contributions to movement-induced hyperemia: implications for rehabilitative medicine. Am J Physiol - Hear Circ Physiol. 2011;300(5):1885–91.
- Haouzi P, Chenuel B, Chalon B. Effects of body position on the ventilatory response following an impulse exercise in humans. J Appl Physiol. 2002;92(4):1423–33.
- Qvarlander S, Sundström N, Malm J, Eklund A. Postural effects on intracranial pressure: modeling and clinical evaluation. J Appl Physiol. 2013;115(10):1474–80.
- Holmlund P, Johansson E, Qvarlander S, Wåhlin A, Ambarki K, Koskinen LOD, et al. Human jugular vein collapse in the upright posture: Implications for postural intracranial pressure regulation. Fluids Barriers CNS. 2017;14(1):1–7.
- Flürenbrock F, Schwander S, Podgoršak A, Bausch B. Insights of posture dependent pressure characteristics in five rats. Curr Dir Biomed Eng. 2021;7(2):799–802.
- Kenny RA, Bayliss J, Ingram A, Sutton R. Head-up tilt: a useful test for investigating unexplained syncope. Lancet. 1986;327:1352–4.
- 19. Sakka L, Chomicki A, Gabrillargues J, Khalil T, Chazal J, Avan P. Validation of a noninvasive test routinely used in otology for the diagnosis of

Page 13 of 14

cerebrospinal fluid shunt malfunction in patients with normal pressure hydrocephalus. J Neurosurg. 2016;124(2):342–9.

- Sutton R, Fedorowski A, Olshansky B, Van Dijk GJ, Abe H, Brignole M, et al. Tilt testing remains a valuable asset. Eur Heart J. 2021;42(17):1654–60.
- Sakka L, Coll G, Chazal J. Anatomy and physiology of cerebrospinal fluid. Eur Ann Otorhinolaryngol Head Neck Dis. 2011;128(6):309–16.
- 22. Bouma GJ, Muizelaar JP, Bandoh K, Marmarou A. Blood pressure and intracranial pressure-volume dynamics in severe head injury: relationship with cerebral blood flow. J Neurosurg. 1992;77(1):15–9.
- Bloomfield GL, Ridings PC, Blocher CR, Marmarou A, Sugerman HJ. A proposed relationship between increased intra-abdominal, intrathoracic, and intracranial pressure. Crit Care Med. 1997;25(3):496–503.
- Podgoršak A, Trimmel NE, Oertel MF, Qvarlander S, Arras M, Eklund A, et al. Intercompartmental communication between the cerebrospinal and adjacent spaces during intrathecal infusions in an acute ovine in vivo model. Fluids Barriers CNS. 2022;19:1–13. https://doi.org/10.1186/ s12987-021-00300-0.
- Flower JO, Parr EA. 13 Control Systems. In: Laughton MA, Warne DJ, editors. Electrical Engineer's Reference Book. Sixteenth. Oxford: Newnes; 2003. p. 13–63.
- Parr EA. 14 Digital control systems. In: Laughton MA, Warne DJ, editors. Electrical Engineer's Reference Book. Sixteenth. Oxford: Newnes; 2003. p. 14–36.
- Wagshul ME, Eide PK, Madsen JR. The pulsating brain: a review of experimental and clinical studies of intracranial pulsatility. Fluids Barriers CNS. 2011;8(1):1–23.
- Zou R, Park E-H, Kelly EM, Egnor M, Wagshul ME, Madsen JR. Intracranial pressure waves: characterization of a pulsation absorber with notch filter properties using systems analysis: Laboratory investigation. J Neurosurg Pediatr PED. 2008;2(1):83–94.
- 29. Animals NRC (US) C for the U of the G for the C and U of L. Guide for the Care and Use of Laboratory Animals. Washington (DC); 2011.
- Pedersen SH, Andresen M, Lilja-Cyron A, Petersen LG, Juhler M. Lumbar puncture position influences intracranial pressure. Acta Neurochir (Wien). 2021;163(7):1997–2004.
- 31. Gustafsson F. Determining the initial states in forward-backward filtering. IEEE Trans Signal Process. 1996;44(4):988–92.
- Trimmel NE, Podgoršak A, Oertel MF, Jucker S, Arras M, Schmid Daners M, et al. The sheep as a comprehensive animal model to investigate interdependent physiological pressure propagation and multiparameter influence on cerebrospinal fluid dynamics. Front Neurosci. 2022. https:// doi.org/10.3389/fnins.2022.868567.
- Trimmel NE, Podgoršak A, Oertel MF, Jucker S, Arras M, Schmid Daners M, et al. Venous dynamics in anesthetized sheep govern postural-induced changes in cerebrospinal fluid pressure comparable to those in humans. Physiol Rep. 2022;10(24): e15525.
- Jardine DL, Wieling W, Brignole M, Lenders JWM, Sutton R, Stewart J. The pathophysiology of the vasovagal response. Hear Rhythm. 2018;15(6):921–9.
- Smith JJ, Porth CM, Erickson M. Hemodynamic response to the upright posture. J Clin Pharmacol. 1994;34(5):375–86. https://doi.org/10.1002/j. 1552-4604.1994.tb04977.x.
- Blomqvist CG, Stone HL. Cardiovascular Adjustments to Gravitational Stress. In: Comprehensive Physiology. Wiley; 2011, 1025–63. https://doi. org/10.1002/cphy.cp020328
- Coonan TJ, Hope CE. Cardio-respiratory effects of change of body position. Can Anaesth Soc J. 1983;30(4):424–37. https://doi.org/10.1007/ BF03007869.
- Blanco CE, Dawes GS, Hanson MA, Mccooke HB. Carotid baroreceptors in fetal and newborn sheep. Pediatr Res. 1988;24(3):342–6.
- Sato M, Tanaka M, Umehara S, Nishikawa T. Baroreflex control of heart rate during and after propofol infusion in humans. Br J Anaesth. 2005;94(5):577–81.
- Assavapokee T, Thadanipon K. Examination of the Neck Veins. N Engl J Med. 2020;383(24): e132. https://doi.org/10.1056/NEJMvcm1806474.
- Gisolf J, van Lieshout JJ, van Heusden K, Pott F, Stok WJ, Karemaker JM. Human cerebral venous outflow pathway depends on posture and central venous pressure. J Physiol. 2004;560(1):317–27.
- Vasquez DG, Berg-Copas GM, Wetta-Hall R. Influence of semi-recumbent position on intra-abdominal pressure as measured by bladder pressure. J Surg Res. 2007;139(2):280–5.

- Cresswell AB, Jassem W, Srinivasan P, Prachalias AA, Sizer E, Burnal W, et al. The effect of body position on compartmental intra-abdominal pressure following liver transplantation. Ann Intensive Care. 2012;2012(Suppl 1):1–10.
- Membrive CMB. Anatomy and Physiology of the Rumen. In: Rumenology. 2016. p. 1–38.
- 45. Czosnyka M, Citerio G. Brain compliance: the old story with a new' et cetera.'Intensive Care Med. 2012;38:925–7.
- Brasil S, Jorge D, Solla F, Nogueira RDC, Teixeira MJ, Marcelo L, et al. Intracranial compliance assessed by intracranial pressure pulse waveform. Brain Sci. 2021;11(8):971.
- Laurent S, Boutouyrie P. Arterial stiffness and hypertension in the elderly. Front Cardiovasc Med. 2020. https://doi.org/10.3389/fcvm.2020.544302.
- 48. Harary M, Dolmans RGF, Gormley WB. Intracranial pressure monitoring review and avenues for development. Sensors. 2018;18:465.
- Wagshul ME, Kelly EJ, Yu HJ, Garlick B, Zimmermann T, Egnor MR. Resonant and notch behavior in intracranial pressure dynamics. J Neurosurg Pediatr. 2009;3(May):354–64.
- Tenti G, Sivaloganathan S, Drake JM. The synchrony of arterial and CSF pulsations is not due to resonance. Pediatr Neurosurg. 2002;37(4):221–2. https://doi.org/10.1159/000065400.
- Shapiro K, Fried A. The theoretical requirements of shunt design as determined by biomechanical testing in pediatric hydrocephalus. Child's Nerv Syst. 1988;4(6):348–53.
- Lee H, Xie L, Yu M, Kang H, Feng T, Deane R, et al. The effect of body posture on brain glymphatic transport. J Neurosci Off J Soc Neurosci. 2015;35(31):11034–44.
- 53. Spencer MP, Denison AB. Pulsatile blood flow in the vascular system. Handb Physiol. 1963;11(2).

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

